



CERTIFICATE OF ACCREDITATION

This is to attest that

TRANSCAL

NO. 100, 10TH CROSS, BETWEEN SAMPIGE ROAD AND MARGOSA ROAD, MALLESWARAM
BANGALORE 560003, REPUBLIC OF INDIA

Calibration Laboratory CL-233

has met the requirements of AC204, *IAS Accreditation Criteria for Calibration Laboratories*, and has demonstrated compliance with ISO/IEC Standard 17025:2017, *General requirements for the competence of testing and calibration laboratories*. This organization is accredited to provide the services specified in the scope of accreditation.

Effective Date August 17, 2021

Expiration Date September 1, 2023



A handwritten signature in black ink that reads 'Raj Nathan'.

President

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

TRANSCAL

Contact Name Saraswathi K

Contact Phone +91-8095111880

Accredited to ISO/IEC 17025:2017

Effective Date August 17, 2021

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)*

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (\pm)	CALIBRATION PROCEDURE AND/OR STANDARD EQUIPMENT USED
<i>Dimensional</i>			
Height Gauge (Vernier/Dial/ Digital)	0 mm to 600 mm 0 mm to 1000 mm	2.2 μ m 2.4 μ m	Using Gauge Blocks/Caliper Checker. IS 2921
V- Block (Parallelism, Symmetry)	300 mm x 125 mm x 200 mm	4.4 μ m	Using Lever Dial Gauge & Mandrel, IS 2949
Caliper (Vernier/Dial/ Digital)	0 mm to 600 mm 0 mm to 1000 mm 0 mm to 2000 mm	7.8 μ m 10 μ m 12 μ m	Using Gauge Block Set, Caliper Checker, IS 3651
External Micrometer (Mechanical / Electronic / Digital)	0 mm to 25 mm	0.4 μ m	Using Gauge Blocks, IS 2967
	0 mm to 150 mm	0.6 μ m	Using Gauge Blocks, IS 2967
	0 mm to 1000 mm 0 mm to 2000 mm	1.7 μ m 11 μ m	
Electronic Probe	0 to 25 mm 0 to 50 mm	0.5 μ m 0.9 μ m	Using Gauge Blocks, By Comparison Method
Wire Gauge	0.19 mm to 7.62 mm	8 μ m	Using Vision System, By Comparison Method
Hegman Gauge	up to 1 mm	2.3 μ m	Using Plunger Dial Gauge By Comparison Method
Wet Film Thickness Gauge	0.025 mm to 5 mm	5.9 μ m	Vision System, By Comparison Method
Depth Micrometer	0 mm to 300 mm	6 μ m	Using Gauge Block Set JIS B 7544
	0 mm to 600 mm	11 μ m	
Depth Caliper	0 mm to 300 mm 0 mm to 600 mm	6.4 μ m 13 μ m	Using Gauge Block Set, IS 4213
Micrometer Setting Rod	25 mm to 1000 mm	4.7 μ m	Using Gauge Block Set, By Comparison Method
	1000 mm to 1950 mm	7.9 μ m	

* If information in this CMC is presented in non-SI units, the conversion factors stated in NIST Special Publication 811 "Guide for the Use of the International System of Units (SI)" apply.

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION PROCEDURE AND/OR STANDARD EQUIPMENT USED
Angle Graticule	0 deg to 360°	1.8'	Using Vision System, Comparison method
Laser Distance Meter	0 mm to 2000 mm	350 µm	Using Slip Gauge, Comparison Method
Inclinometer	0° to 90°	1.7 mins of Arc	Using Angle Gauge Blocks, Comparison Method based on IS 4239
Bench Center (Co-axiality)	0 mm to 500 mm	8.9 µm	Using Straight & Taper Mandrels, Dial Gauge. IS 5980
Angle Plate	450 mm x 300 mm x 200 mm	11 µm	Using Liver Dial Gauge, IS 2554, IS 6973
Flakiness Gauge	0 mm to 100 mm	5.9 µm	Using Vision System & 2D Height Gauge, Comparison Method
Elongation Gauge	0 mm to 100 mm	5.9 µm	Using 2D Height Gauge, Comparison Method
Dial Gauge (Plunger /Digital/ Dial Thickness Gauge)	0 mm to 100 mm	0.7 µm	Using Length Measuring Machine, IS 2092
Dial Gauge (Lever Type)	0 mm to 2 mm	0.5 µm	Using Length Measuring Machine, IS 11498
Bore Dial Gauge for Transmission Accuracy check	0 mm to 2 mm	1.8 µm	Using Length Measuring Machine, LIS B 7515
Surface Plate	5 m X 3 m	$1.3 \sqrt{((L+W)/100)}$ L = Length in mm W = Width in mm	Using Electronic Level, IS 7327, IS 2285
Plain Plug Gauge	0 mm to 100 mm 100 mm to 400 mm	1 µm 1.6 µm	Using Length Measuring Machine, Master Disc, FCDM, IS 3455
Feeler Gauge	Up to 1 mm	1.4 µm	Using Digital Micrometer, IS 3179
Cylindrical Measuring Pin	0.1 mm to 26 mm	0.9 µm	Using Length Measuring Machine, Master Disc, FCDM, IS 1103
Thread Plug Gauge (Major Diameter & Effective Diameter)	1 mm to 3 mm 3 mm to 100 mm 100 mm to 400 mm	1.2 µm 1.2 µm 1.7 µm	Using Length Measuring Machine, Master Disc, FCDM, IS 2334
Snap Gauge	3 mm to 500 mm	2.7 µm	Using Gauge Block Set, IS 3477

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION PROCEDURE AND/OR STANDARD EQUIPMENT USED
Bevel Protractor / Combination Set	0° to 360°	2.9'	Using Angle Block Set, IS 4239
Thread Ring Gauge (For Effective Diameter & Minor Diameter only)	3 mm to 325 mm	1.9 µm	Using Length Measuring Machine, Master Ring, IS 2334, IS 4218
Spirit Level (Type 1, 2 & 3)	Sensitivity up to 0.01 mm/m	0.003 mm/m	Using Electronic probe with DRO / Electronic Level, IS 5706 & IS 1632,
Plain Ring Gauge	3 mm to 325 mm	1.9 µm	Using Length Measuring Machine, Master Ring, IS 3455
Measuring Scales	Up to 3000 mm	114 \sqrt{L} µm where L is length in m	Using Length Measuring Machine, IS 1481
Measuring Tape/ Pi Tape	Up to 50 m	114 \sqrt{L} µm where L is length in m	Using Length Measuring Machine, IS 1269
Thread Pitch Gauge Pitch	0.25 mm to 6.35 mm	4.8 µm	Using Vision System, IS 4211
Thread Pitch Gauge Angle	55° 60°	1.2' 1.2'	Using Vision System, IS 4211
Radius Gauge	0.4 mm to 50 mm	5.6 µm	Using Vision System, IS 5273
Internal / Stick Micrometer (2 Point)	50 mm to 5000 mm	(1.25 + 3.27L) µm where L is length in m	Using Gauge Block Set, IS 2966
Engineering Square (Squareness)	Up to 400 mm	6.2 µm	Using Granite Square & Slip Gauge, IS 2103
Test Sieves	0.03 mm to 125 mm	4.8 µm	Using Vision System, IS 460
Comparator Stand (Flatness of worktable)	200 mm x 200 mm	2.4 µm	Using Lever Dial Gauge, IS 7599
Straight Edge/Parallels	Up to 2000 mm	2.8 µm	Using Electronic Level, IS 2220
Inside/ Outside Dial Caliper	0 mm to 150 mm	0.7 µm	Using Gauge Block Set, By Comparison Method.
Pistol Caliper	0 mm to 100 mm	60 µm	Using Gauge Block Set, By Comparison Method.
Thickness Plate/Foils	Up to 2.5 mm	1.6 µm	Using Digital Micrometer, By Comparison Method.
Limit Gauge /Test probes (Length, Radius / Diameter, Angle)	0 mm to 400 mm 360°	4.6 µm 2.4'	Using Vision System, By Comparison Method.
Coating Thickness Gauge with Foils	0 mm to 2000 µm	1.8 µm	Using Standard Thickness Foils, By Comparison Method.
Taper Thread Plug Gauge (Up to 150 mm diameter)	Taper angle and Flank angle	3.2"	Using Length Measuring Machine, Master Disc, FCDM; Vision system, IS 8999
	Diameter at ends	1.4 µm	

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION PROCEDURE AND/OR STANDARD EQUIPMENT USED
3-Point Micrometer	2.5 mm to 100 mm	2.0 µm	Using set of Setting Ring Gauges, By Comparison Method.
Taper Scale	1 mm to 15 mm	4.8 µm	Using Vision System, By Comparison Method.
Ultrasonic Thickness Gauge	0 mm to 300 mm	38 µm	Using Gauge Block Set, By Comparison Method.
Length Bars	50 mm to 500 mm	2.9 µm	Using Length Measuring Machine, Based on IS 2984
Micrometer Head	0 mm to 50 mm	0.8 µm	Using Length Measuring Machine, IS 2966
Fillet Gauge/ Form Gauge	0 mm to 150 mm 0° to 90°	5.8 µm 5.4'	Using Vision System, By Comparison Method.
Taper Plain Plug Gauge	Taper Half Angle Up to 100 mm	0.6" 0.9 µm	Using Length Measuring Machine, IS 9475, IS 2251
Taper Thread Ring Gauge	Up to 100 mm	0.9 µm	Using Length Measuring Machine, IS 8999, ASME B1.20.5
Taper Plain Ring Gauge	Taper Half Angle Up to 100 mm	5.5" 0.9 µm	Using Length Measuring machine, IS 9475, IS 2251
Sine Bar/ Sine Centre / Sine Table	0° to 45°	2.8"	Using Gauge Blocks, Angle Blocks, Lever Dial Gauge, IS 5359
Profile Projector / Video Measuring machine /Microscope	10 X to 50 X	0.06 %	Using Gauge Blocks/Glass Scale, JIS B7184
	0° to 360°	2.8"	Using Angle Gauge Blocks/Glass Scale, JIS B7184
	400 mm	2.9 µm	Using Gauge Blocks/Glass Scale, JIS B7184
Floating Carriage Micrometer	0 mm to 100 mm	1.5 µm	Using Mandrels & Master Cylinders, MOY/SCMI/9
Thread Measuring Wire	0.170 mm to 6.350 mm	0.5 µm	Using Electronic Probe with DRO, Comparison Method
Cylindrical Master (Diameter & Concentricity)	3 mm to 100 mm	1.1 µm (diameter) 1.3 µm (concentricity)	Electronic Probe with DRO, IS 6311
Slip Gauges (Gauge Blocks) / Micrometer Check set	0.5 mm to 25 mm 25 mm to 50 mm 50 mm to 75 mm 75 mm to 100 mm	0.07 µm 0.10 µm 0.12 µm 0.15 µm	Slip Gauge Calibrator & k Grade Slip Gauges, IS 2984, ISO 3650
Dial Calibration Tester	0 mm to 25 mm	0.7 µm	Using Electronic Probe with DRO, by Comparison Method

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION PROCEDURE AND/OR STANDARD EQUIPMENT USED
Gauge Block Comparators	0 mm to 100 mm	0.03 µm	Using K grade Gauge Blocks by Direct Method
Surface Roughness Master specimen	Ra up to 10 µm Rz up to 25 µm	8 % 8 %	Using Surface Roughness Tester, IS 3073, IS 10707
Surface Roughness Tester	1 µm to 800 µm	8.1 %	Using Depth Master and roughness master, By Direct Method
Caliper Checker	20 mm to 600 mm 20mm to 1000 mm	2.1 µm 3.6 µm	Electronic Comparator and gauge block (0 grade) Using CMM, By Direct Method
Length Measuring Machine	0 mm to 100 mm	0.3 µm	Using Gauge Block Set (0 Grade), Comparison Method
	0 mm to 1000 mm (Tape & Scale Calibration)	3.7 µm	Using Gauge Block Set and Long Gauge Block (0 Grade), Comparison Method
Long Gauge Blocks	100 mm to 300 mm 300 mm to 500 mm	1.6 µm 1.8 µm	Electronic Comparator and gauge blocks (K Grade), IS 2984
Glass Scale	0 mm to 300 mm	1.5 µm	Using Vision Measuring System, Comparison Method
Granite square	Up to 600 mm x 600 mm x 100 mm Flatness Squareness		Using CMM, Based On IS:2103
		2.5 µm 2.8 µm	
Taper / conical/ Cylindrical Mandrel	Major/Cylindrical diameter 3 mm to 200 mm Taper angle Gauge plane Ø Roundness Straightness Run out		Using CMM, Based On IS:2063, ISO 230 - 1
		2"	
		2.8 µm	
		2.8 µm	
		2.8 µm	
		3.0 µm	
Jigs, Fixtures, PCD gauges, lever arm, master Block	Up to 500 X 800 mm Linear Measurement Angle	3.6 µm 1.6"	Using CMM, By Direct Method
Coordinate Measuring Machine	Up to 1000 mm X,Y,Z total measuring error (MPEE) Probing Error	(1.3+0.2L) µm, where L in meters	Using Gauge block set grade '0', Step Gauge Block Grade '0' IS:15635, ISO 10360
		1.2 µm	

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (\pm)	CALIBRATION PROCEDURE AND/OR STANDARD EQUIPMENT USED
Optical Flat and Parallels	Diameter up to 100 mm Thickness up to 25 mm Flatness Parallelity Size	0.07 μ m 0.28 μ m 0.35 μ m	Using Master flat with Monochromatic light source, Gauge blocks grade '0' & Electronic Comparator, IS:5440
Angle gauges	0° to 90°	6"	Using Sine bar, Gauge blocks grade '0' and Electronic Comparator, by comparison Method
Contour Tracer Machine	\pm 25 mm Linear Measurement Angle Radius	1.8 μ m 36" 1.3 μ m	Using Gauge blocks grade '0', Angle gauge & Cylindrical Master, by comparison Method
Gauge Block Accessories	up to 600mm Flatness Parallelism Size	0.23 μ m 0.28 μ m 1.8 μ m	Using Optical flat, Gauge blocks, electronic comparator, IS 4440.
Sphere/Spherical Ball Diameter	Up to 100 mm	0.4 μ m,	Using LMM, By Direct Method
Electronic Level	\pm 2 mm/m	1.7 μ m/m	Electronic probe with DRO with tilting fixture JIS B 7510
Spline gauge (for Major, Minor & Over pins, Diameters only)	10 mm to 100 mm	2.1 μ m	Using VMS/FCDM, By Direct Method, IS 2327/ISO14
Electronic Comparator (Lever Type)	up to 2 mm	0.3 μ m	Using LMM, By Comparison Method
Mechanical			
Sound Level Meter / Sound Level Calibrator (at 1 kHz)	94 dB and 114 dB	0.3 dB	Using Sound Level Calibrator / Sound Level Meter, By Comparison Method
Vibration Machine/Vibration Measurement (5 Hz to 10 kHz)	0.1 g to 40 g (where g is acceleration due to gravity)	2 %	Using Accelerometer and Digital Oscilloscope, By Comparison Method
Accelerometer	5 Hz to 100 Hz 101 Hz to 500 Hz 501 Hz to 5000 Hz 5001 Hz to 10000 Hz	2.6 % 2.7 % 3.2 % 4.1 %	Using Portable Vibration Calibrator as per ISO 16063 - 21
Portable Vibration Calibrator	5 Hz to 9 Hz 10 Hz to 99 Hz 100 Hz to 1999 Hz 2000 Hz to 10000 Hz	2.1 % 1.6 % 1.4 % 2.6 %	Using Accelerometer and Digital Oscilloscope as per ISO 16063 - 21

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION PROCEDURE AND/OR STANDARD EQUIPMENT USED
Vibration Meter	Acceleration: 1 m/s ² to 200 m/s ² (10 Hz to 1000 Hz)	3.2 %	Using Portable Vibration Calibrator as per ISO 16063 - 21
	Velocity: 1 mm/s to 200 mm/s (10 Hz to 1000Hz)	2.7 %	
	Displacement: 0.01 mm to 4 mm (10 Hz to 100 Hz)	2.6 %	
Volumetric Glassware (One Mark Pipette, Graduated Pipette, Graduated Burette, Volumetric Flask, Measuring Jar)	1 mL to 10 mL 11 mL to 100 mL 101 mL to 1000 mL 1001 mL to 5000 mL 5001 mL to 20000 mL	0.00018 mL 0.0004 mL 0.005 mL 0.018 mL 0.7 mL	Calibration based on Gravimetric method as per ISO 4787 using weighing balance of d = 0.1 mg, 1 mg, 10,100 mg and Distilled Water
Micro Pipette	1 µL to 10 µL 10 µL to 100 µL 100 µL to 1000 µL 1000 µL to 10000 µL	0.02 µL 0.05 µL 0.11 µL 1.9 µL	Calibration based on Gravimetric method as per ISO 8655 using weighing balance of d = 0.001 mg, 0.01 mg and Distilled Water
Pressure Measurement Devices (Digital / Analogue Pressure Gauges, Differential Pressure Gauge, Transducers, Transmitters, Switches)	0.1 mbar to 10 mbar 10 mbar to 100 mbar 100 mbar to 2 bar 2 bar to 40 bar 40 bar to 700 bar 700 bar to 1000 bar	0.35 % 0.06 % 0.042 % 0.026 % 0.023 % 0.021	Digital Pressure Gauge using Pneumatic / Hydraulic Comparator Pump based on DKD – R 6 - 1
	2 bar to 20 bar	0.0066 %	Digital Pressure Controller using Pneumatic Comparator Pump based on DKD – R 6- 1
	1 bar to 35 bar 35 bar to 1200 bar	0.015 % 0.016 %	Using Hydraulic (Oil operated) Dead Weight Tester based on DKD – R6 - 1
Dead Weight Tester (Hydraulic)	6 bar to 60 bar 60 bar to 1200 bar	0.01 % 0.0093 %	Using Dead Weight Tester as per Euramet cg-3 & OIML R 110 by Cross Float method / Effective Area and Pressure method
Pirani gauges/ Penning Gauges	0.001 µbar to 0.001 mbar 0.001 mbar to 0.06 mbar 0.06 mbar to 5 mbar 5 mbar to 100 mbar 100 mbar to 1000 mbar	3.5 % 2.5 % 1.5 % 1.2 % 1 %	Using Standard Pirani Gauge / Penning Gauge, ISO19685

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION PROCEDURE AND/OR STANDARD EQUIPMENT USED
Vacuum Measurement Devices (Digital/Analogue Vacuum Gauges, Transducers/ Transmitters)	-0.01 bar to -0.9 bar	0.017 %	Digital Pressure Gauge using Pneumatic Vacuum Comparator Pump based on DKD – R6 - 1
Absolute Pressure Gauges	100 mbar to 915 mbar (abs)	0.26 %	Standard absolute Gauge using desiccator and vacuum pump, OIML R97
Altimeter Chamber	30 mbar to 1000 mbar	3 mbar	Using Barometer, By Comparison Method
Calibration of E1 Class Weights and coarser	1 mg	0.0010 mg	Using E1 class Standard weights and Mass comparator of $d = 0.1 \mu\text{g}$, per OIML R – 111 by subdivision through ABBA cycles
	2 mg	0.0010 mg	
	5 mg	0.0010 mg	
	10 mg	0.0010 mg	Using Mass Comparator of $d = 0.001 \text{ mg}$, per OIML R 111 by Substitution Method
	20 mg	0.0010 mg	
	50 mg	0.0010 mg	
	100 mg	0.0013 mg	
	200 mg	0.0014 mg	
	500 mg	0.0016 mg	
	1 g	0.003 mg	Using Mass Comparator of $d = 0.01 \text{ mg}$, per OIML R 111 by Substitution Method
	2 g	0.004 mg	
	5 g	0.005 mg	
	10 g	0.006 mg	
	20 g	0.007 mg	
	50 g	0.01 mg	
	100 g	0.02 mg	Using Mass Comparator of $d = 100 \text{ mg}$, by Substitution Methods through ABBA Cycles as per OIML R - 111
	200 g	0.03 mg	
	500 g	0.10 mg	
	1 kg	0.20 mg	
Calibration of E2 Class Weights and coarser	2 kg	1 mg	Using Mass Comparator of $d = 1 \text{ mg}$, by Substitution Methods through ABBA Cycles as per OIML R - 111
	5 kg	2 mg	
	10 kg	3 mg	Using Mass Comparator of $d = 100 \text{ mg}$, by Substitution Methods through ABBA Cycles as per OIML R - 111
	20 kg	7 mg	
	50 kg	100 mg	Using Mass Comparator of $d = 100 \text{ mg}$, by Substitution Methods through ABBA Cycles as per OIML R - 111
Calibration of Class 1 Weighing Balances and coarser	1 mg to 2 g	0.0035 mg	E1 Class Standard Weights 1 mg to 20 kg, as per OIML R - 76
	1 mg to 5 g	0.005 mg	
	1 mg to 20 g	0.007 mg	
	1 mg to 50 g	0.02 mg	
	1 mg to 200 g	0.03 mg	
	1 mg to 500 g	0.10 mg	
	1 mg to 1 kg	0.20 mg	
	500 mg to 5 kg	1 mg	
	500 mg to 20 kg	7 mg	

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION PROCEDURE AND/OR STANDARD EQUIPMENT USED
Calibration of Class 2 Weighing Balances and coarser	500 mg to 50 kg 500 mg to 150 g 1 kg to 300 kg 1 kg to 1000 kg 2 kg to 3000 kg	100 mg 1 g 10 g 100 g 500 g	F1 and M1 class Standard weights Up to 3000 kg, as per OIML R - 76
Spring Balance	10 g to 1500 g 1500 g to 100 kg	0.28 % 0.1 %	Using F1 class weights, By Comparison Method
Torque Wrench, Torque Driver Type I - Class B, C, D, E Type II - Class A, B, D, E	0.1 N·m to 1 N·m 1 N m to 10 N m 10 N·m to 20 N·m 20 N·m to 200 N·m 200 N·m to 1000 N·m	0.89 % 0.63 % 0.38 % 0.43 % 0.47 %	Torque sensors of various capacities using Torque Calibration Rig Based on ISO 6789
Torque Measuring Devices	0.01 N·m to 0.1 N·m 0.1 N·m to 1 N·m 1 N·m to 10 N·m 10 N·m to 500 N·m	0.1 % 0.03 % 0.03 % 0.03 %	Using Dead Weight Torque Calibration System as per BS 7882
Load Cells / Force Proving Instruments	0.1 N to 100 N 100 N to 2000 N	0.07 % 0.11 %	Using Newton weights as per ISO 376
Universal Testing Machine (UTM)	Compression: 20 N to 10 kN 20 kN to 1000 kN	0.52 % 0.52 %	Using Force Proving Instruments as per IS 1828
	Tension: 20 N to 10 kN 20 kN to 100 kN	0.37 % 0.37 %	
Displacement Measuring System and Devices used in Material Testing Machine	0 mm to 5 mm 0 mm to 75 mm 0 mm to 500 mm 0 mm to 1000 mm	0.003 mm 0.030 mm 0.035 mm 0.088 mm	Using Gauge Blocks as per ASTM E2309
Extensometer used in Material Testing Machine - Gauge Length up to 200 mm	0 mm to 3 mm 0 mm to 10 mm 0 mm to 12.5 mm 0 mm to 25 mm	0.010 mm 0.030 mm 0.055 mm 0.087 mm	Using Digital Micrometer as per ISO 9513 and ASTM E83
Speed of Material Testing Machine	0 mm/min to 5 mm/min 0 mm/min to 50 mm/min 0 mm/min to 100 mm/min 0 mm/min to 500 mm/min 0 mm/min to 800 mm/min	0.06 mm/min 0.06 mm/min 0.12 mm/min 0.58 mm/min 0.93 mm/min	Using Stop watch as per ASTM E2658
Push Pull Gauge / Gram Gauge / Test Fingernail Probe/Tension Gauge	1 N to 2000 N	0.21 %	Newtonian Weights And Frame Fixture VDI/VDE 2624 – Part 2.1
Specific Gravity Hydrometer/ Density / Baume / Brix Hydrometer / Lactometer / Alcoholmeter	SG = 0.6 to 1.0 SG = 1.0 to 2.0	0.00014 0.00014	Calibration of Hydrometers by Cuckows method

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION PROCEDURE AND/OR STANDARD EQUIPMENT USED
Viscosity (Viscosity Cups, Zahn Cups)	30 cSt to 240 cSt	0.33 %	Calibration as per IS 3944, ASTM D4212 using liquids of known kinematic viscosity
Liquid Flow Meter	1 m ³ /h to 350 m ³ /h	0.54 %	Using Clamp-On Ultra Sonic Flow Meter by Comparison Method
Flow Meters (Air)	0.1 L/min to 5 L/min 5 L/min to 500 L/min	0.52 % 0.27 %	Using Mass Flow Meter by Comparison Method
Air Velocity (Anemometers, biosafety cabinets, thermionics instruments)	0.3 m/s to 30 m/s	1.3 %	Using Digital Anemometer by comparison method or direct method
Liquid Flow Measuring Devices	5 mL/h to 200 L/h 200 L/h to 4500 L/h	0.11 % 0.17 %	Using Weighing System by Gravimetric Method
Rubber Hardness Tester	Shore A, B, E, O Shore C, D, DO Shore M, OO, OOO, OOO-S	0.1 % 0.1 % 0.057 Shore	Using Load Cell with Indicator as per ASTM D 2240 Using Electronic Balance as per ASTM D 2240
Rockwell Hardness Testing Machines	HRA HRBW HRC	0.56 HRA 0.59 HRBW 0.56 HRC	Using Standard Hardness Blocks as per IS 1586 (Part II) / ISO 6508 - 2 by Indirect Method
Brinell Hardness Testing Machines	HBW 2.5 / 187.5 HBW 5 / 750 HBW 10 / 3000	1 % 1 % 1 %	Using Standard Hardness Blocks as per IS 1500 (Part II) / ISO 6506 - 2 by Indirect Method
Vickers Hardness Testing Machines	HV 10 HV 20 HV 30 HV 50	1.1 % 1.1 % 1.1 % 1.1 %	Using Standard Hardness Blocks as per IS 1501 (Part II) / ISO 6507 - 2 by Indirect Method
Thermal			
Humidity Meters (Dial / Digital)	0.5 %RH 5 %RH 95 %RH	0.33 %RH 0.33 %RH 0.85 %RH	Using Humidity standard solution by direct method
Dial /Digital Humidity Meters	10 %RH to 95 %RH (at 10 °C to 60 °C)	0.80 %RH	Using Temperature & Humidity Meter with Humidity Chamber by Comparison Method
Humidity Transmitters (for temperature scale)	0 °C to 60 °C	0.18 °C	Using Class 'A' RTD Sensor /PRT sensor with Digital Indicator by comparison
RTD, Thermocouples, Indicator with sensor	-196 °C	0.08 °C	Using LN ₂ and cryo bath by comparison method

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION PROCEDURE AND/OR STANDARD EQUIPMENT USED
RTD, Thermocouples, Indicator with sensor continued	-100 °C to -45 °C -45 °C to 140 °C 140 °C to 650 °C	0.09 °C 0.02 °C 0.08 °C	Using Dry Temperature Bath, SPRT with Digital Indicator by comparison Method
	650 °C to 1000 °C 1000 °C to 1200 °C	0.43 °C 1.5 °C	Using Dry Temperature Bath, S- Type Thermo couple with Digital Indicator by Comparison Method
Glass Thermometer	-80 °C to 250 °C	0.18 °C	Using Liquid Bath, SPRT with Digital Indicator by Comparison Method
Oil Bath, Low & High Temperature Bath, Dry Bath, Incubators & Autoclave, Thermal Chambers / Ovens, Water Bath, Furnace	-100 °C to 140 °C 140 °C to 650 °C	0.072 °C 0.08 °C	Using SPRT with Digital Indicator by Direct Method
	650 °C to 1200 °C	1.5 °C	Using S-Type Thermo couple with Digital Indicator by Direct Method
Humidity Chambers, Climatic Chambers, Dry cabinet, De-humidifier	10 %RH to 95 %RH (at 10 °C to 60 °C)	0.98 %RH	Using Temperature & Humidity Meter by Direct Method
	5 %RH to 10 %RH @ ambient temperature	1.3 %RH	
IR Thermometer	-20 °C to 0 °C	0.45 °C	By Comparison Method
	0 °C to 50 °C	1.0 °C	
	50 °C to 650 °C	1.0 °C	
	650 °C to 1200 °C	1.7 °C	
Electrical – DC/LF			
AC Voltage Generate ³	50 µV to 2 mV (50 Hz to 1 kHz)	1 %	Using Calibrator Fluke 5700A/5522A by Direct Method
	2 mV to 20 mV (10 Hz to 40 Hz)	700 µV/V + 5 µV	
	(40 Hz to 20 kHz)	190 µV/V + 5 µV	
	(20 kHz to 50 kHz)	460 µV/V + 5.6 µV	
	(50 kHz to 100 kHz)	850 µV/V + 7 µV	
	(100 kHz to 300 kHz)	0.12 % + 16 µV	
(300 kHz to 500 kHz)	0.17 % + 25 µV		
(500 kHz to 1 MHz)	0.45 % + 5 µV		
20 mV to 200 mV (10 Hz to 40 Hz)	320 µV/V + 8 µV		
(40 Hz to 20 kHz)	140 µV/V + 9 µV		
(20 kHz to 50 kHz)	380 µV/V + 9 µV		
20 mV to 200 mV (50 kHz to 300 kHz)	0.13 % + 30 µV		
(300 kHz to 1 MHz)	0.41 % + 100 µV		
200 mV to 2 V			

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (\pm)	CALIBRATION PROCEDURE AND/OR STANDARD EQUIPMENT USED
AC Voltage Generate ³ continued	(10 Hz to 40 Hz) (40 Hz to 20 kHz) (20 kHz to 50 kHz) (50 kHz to 300 kHz) (300 kHz to 1 MHz)	175 μ V/V + 35 μ V 95 μ V/V + 10 μ V 145 μ V/V + 20 μ V 0.04 % + 70 μ V 0.3 % + 1 mV	Using Calibrator Fluke 5700A/5522A by Direct Method
	2 V to 20 V (10 Hz to 40 Hz) (40 Hz to 20 kHz) (20 kHz to 50 kHz) (50 kHz to 300 kHz) (300 kHz to 1 MHz)	175 μ V/V + 350 μ V 95 μ V/V + 100 μ V 145 μ V/V + 200 μ V 0.04 % + 700 μ V 0.3 % + 12 mV	
	20 V to 30 V (300 kHz to 500 kHz)	0.9 % + 15 mV	
	20 V to 220 V (10 Hz to 40 Hz) (40 Hz to 20 kHz) (20 kHz to 50 kHz) (50 kHz to 100 kHz)	175 μ V/V + 3.5 mV 100 μ V/V + 1.1 mV 240 μ V/V + 3.5 mV 0.06 % + 10 mV	
	200 V to 1000 V (50 Hz to 1 kHz)	150 μ V/V + 4 mV	
AC Current Generate ³	10 μ A to 200 μ A (40 Hz to 1 kHz)	190 μ A/A + 16 nA	Using Calibrator Fluke 5700A/5522A by Direct Method
	200 μ A to 2 mA (10 Hz to 40 Hz) (40 Hz to 1 kHz) (1 kHz to 5 kHz) (5 kHz to 10 kHz)	650 μ A/A + 35 nA 190 μ A/A + 35 nA 700 μ A/A + 0.4 μ A 0.16 % + 0.8 μ A	
	2 mA to 20 mA (10 Hz to 40 Hz) (40 Hz to 1 kHz) (1 kHz to 5 kHz) (5 kHz to 10 kHz)	650 μ A/A + 350 nA 190 μ A/A + 350 nA 700 μ A/A + 4 μ A 0.16 % + 8 μ A	
	20 mA to 200 mA (10 Hz to 40 Hz) (40 Hz to 1 kHz) (1 kHz to 5 kHz) 20 mA to 200 mA (5 kHz to 10 kHz)	650 μ A/A + 3.5 μ A 190 μ A/A + 3.5 μ A 700 μ A/A + 40 μ A 0.16 % + 80 μ A	
	200 mA to 2 A (40 Hz to 1 kHz)	650 μ A/A + 35 μ A	

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION PROCEDURE AND/OR STANDARD EQUIPMENT USED	
AC Current Generate ³ continued	(1 kHz to 5 kHz) (5 kHz to 10 kHz)	900 µA/A + 80 µA 0.85 % + 160 µA	Using Calibrator Fluke 5700A/5522A by Direct Method	
	2 A to 10 A (45 Hz to 100 Hz) (100 Hz to 1 kHz) (1 kHz to 5 kHz)	0.08 % + 2 mA 0.08 % + 2 mA 3.4 % + 2 mA		
	10 A to 20 A (45 Hz to 100 Hz) (100 Hz to 1 kHz) (1 kHz to 5 kHz)	0.15 % + 5 mA 0.18 % + 5 mA 3.4 % + 5 mA		
	20 A to 120 A (50 Hz to 1 kHz)	0.1 %		
	120 A to 1000 A (50 Hz to 400 Hz)	0.4 %		
	1000 A to 3000 A (50 Hz)	1.2 %		
AC Power Generate ³ Single Phase, 50 Hz, UPF 120 V to 1000 V 0.01 A to 1000 A	0.01 W to 4.8 kW 4.8 kW to 1 MW	0.12 % 0.8 %	Using Calibrator Fluke 5522A with 50 turns current coil, Direct Method	
	AC Power Generate ³ Single Phase, 50 Hz, 0.2 PF lead/lag, 120 V to 1000 V 0.1 A to 1000 A	2.4 W to 200 kW		1 %
	AC Power Generate ³ Single Phase, 50Hz, 0.5 PF lead/lag, 120 V to 1000 V 0.1 A to 1000 A	6 W to 500 kW		0.50 %
AC Power Generate ⁴ Single Phase, 50Hz, 0.8 PF lead/lag, 120 V to 1000 V, 0.1 A to 1000 A	9.6 W to 800 kW	0.23 %		
AC Energy Generate ³ Active /Reactive Single & Three Phase, 40 V to 300 V, 0.05 A to 20 A, 40 Hz to 70 Hz	0.5 Wh to 6 kWh (0.25 PF to 1 PF)	0.25 %	Using Three Phase Energy Source Direct Method	

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (\pm)	CALIBRATION PROCEDURE AND/OR STANDARD EQUIPMENT USED
Power Factor, Three phase	0.25 PF Lag to Unity 0.25 PF Lead to Unity	0.006 PF	Using Edutech Energy source by direct method
Power Factor, Single phase	0.2 PF Lead to unity 0.2 PF Lag to Unity	0.002 PF	Using Calibrator Fluke 5522A by Direct Method
DC Voltage Generate ³	1.018 V, 10 V	3 μ V/V	Using 732B Reference Standard by Direct Method
	50 μ V to 0.5 mV 0.5 mV to 220 mV 220 mV to 2.2 V 2.2 V to 11 V 11 V to 22 V 22V to 220 V 220 V to 1000 V	0.2 % 7 μ V/V + 1 μ V 8 μ V/V + 1 μ V 8 μ V/V + 3.5 μ V 8 μ V/V + 6.5 μ V 9 μ V/V + 80 μ V 11 μ V/V + 500 μ V	Using Calibrator Fluke 5700A by Direct Method
DC Current Generate ³ / Measure ⁴	10 nA to 10 μ A	0.15 %	Using Reference 732B, Decade Megohm box, DMM by VI method
DC Current Generate ³	10 μ A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A 2 A to 10 A 10 A to 20 A	70 μ A/A + 8 nA 62 μ A/A + 8 nA 72 μ A/A + 80 nA 82 μ A/A + 0.8 μ A 100 μ A/A + 25 μ A 0.06 % + 500 μ A 0.12 % +750 μ A	Using Fluke 5700A/5522A, by Direct Method
	20 A to 1000 A 1000 A to 3000 A	0.4 % 1.2 %	Using Current Source with current coil
DC Power Generate ³ (1 V to 1000 V, 1 mA to 1000 A)	1 mW to 1 kW 1 kW to 1 MW	0.08 % 0.65 %	Using Calibrator Fluke 5522A with 50 turns current coil, Direct Method
Discrete DC Resistance Generate ³	1 Ohm 1.9 Ohm 10 Ohm 19 Ohm 100 Ohm 190 Ohm 1 kOhm 1.9 kOhm 10 kOhm 19 kOhm 100 kOhm 190 kOhm 1 MOhm 1.9 MOhm 10 MOhm 19 Mohm 100 MOhm	110 μ Ω / Ω 59 μ Ω / Ω 28 μ Ω / Ω 32 μ Ω / Ω 20 μ Ω / Ω 29 μ Ω / Ω 15 μ Ω / Ω 16 μ Ω / Ω 14 μ Ω / Ω 14 μ Ω / Ω 16 μ Ω / Ω 16 μ Ω / Ω 23 μ Ω / Ω 24 μ Ω / Ω 47 μ Ω / Ω 55 μ Ω / Ω 140 μ Ω / Ω	Using Calibrator Fluke 5700 with DMM 3458A by Direct Method

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION PROCEDURE AND/OR STANDARD EQUIPMENT USED
DC Resistance Generate ³	75 μΩ 0.001 Ω to 0.1 Ω 0.1 Ω to 1 Ω	1 % 0.6 % 0.1 %	Using Standard Resistors & Shunts by V-I Method
	1 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 kΩ 1 kΩ to 10 kΩ 10 kΩ to 100 kΩ 100 kΩ to 1 MΩ 1 MΩ to 10 MΩ 10 MΩ to 100 MΩ 100 MΩ to 300 MΩ 300 MΩ to 1 GΩ	35 uΩ/Ω + 1.2 mΩ 35 uΩ/Ω + 1.4 mΩ 35 uΩ/Ω + 2 mΩ 35 uΩ/Ω + 0.02 Ω 35 uΩ/Ω + 0.2 Ω 40 uΩ/Ω + 2 Ω 150 uΩ/Ω + 50 Ω 600 uΩ/Ω + 3 kΩ 0.36 % + 100 kΩ 1.7 % + 500 kΩ	Using Calibrator Fluke 5522 A by Direct Method
Temperature Simulation – Generate ³ (temperature indicator / controller / recorder)	1 GΩ to 1 TΩ	1 %	Using Decade MegOhm Box
	B Type: 600 °C to 1800 °C K Type: -200 °C to 1372 °C E Type: -200 °C to 1000 °C U Type: -200 °C to 400 °C L Type: -200 °C to 900 °C J Type: -200 °C to 1200 °C T Type: -200 °C to 400 °C N Type: -200 °C to 1300 °C R Type: 0 °C to 1750 °C S Type: 0 °C to 1750 °C	0.5 °C 0.06 °C 0.08 °C 0.07 °C 0.08 °C 0.06 °C 0.1 °C 0.07 °C 0.07 °C 0.07 °C	Using 5700A Calibrator DC mV measurement method
Capacitance Generate ³	RTD: -200 °C to 800 °C	0.07 °C	Using Resistance Method
	(1 kHz) 220 pF to 1 μF 1 μF to 10 μF (100 Hz) 10 μF to 100 μF 100 μF to 1 mF 1 mF to 10 mF 10 mF to 110 mF	0.05 % + 10 pF 0.25 % + 15 nF 0.25 % + 150 nF 0.45 % + 1 μF 0.45 % + 10 μF 1.3 %	Using Calibrator Fluke 5520A, Decade Capacitance Box by Direct Method
Harmonics Generate ³ (45 Hz to 5 kHz)	2nd order to 39th order (33 mV to 1000 V and 3.3 mA to 20 A)	0.5 %	FLUKE 5522A Source, Direct Method
AC Voltage Measure ⁴	1 mV to 200 mV (50 Hz to 2 kHz)	130 uV/V + 4 uV	Using 8½ DMM 8508A/HP 3458 A, by Direct Method
	100 mV to 2 V (20 Hz to 50 Hz) (2 kHz to 100 kHz)	130 uV/V + 20 uV 600 uV/V + 0.2 mV	

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION PROCEDURE AND/OR STANDARD EQUIPMENT USED
AC Voltage Measure ⁴ (continued)	200 mV to 2 V (50 Hz to 2 kHz)	95 μ V/V + 20 μ V	Using 8½ DMM 8508A/HP 3458 A, by Direct Method
	1 V to 20 V (100 kHz to 1 MHz)	1 % + 0.2 V	
	2 V to 20 V (20 Hz to 50 Hz) (50 Hz to 2 kHz) (2 kHz to 100 kHz)	130 μ V/V + 0.2 mV 95 μ V/V + 0.2 mV 600 μ V/V + 2 mV	
	20 V to 200 V (20 Hz to 50 Hz) (50 Hz to 2 kHz) (2 kHz to 100 kHz)	130 μ V/V + 2 mV 95 μ V/V + 2 mV 600 μ V/V + 20 mV	
	100 V to 1000 V (50 Hz to 10 kHz) (10 kHz to 20 kHz)	130 μ V/V + 20 mV 270 μ V/V + 40 mV	
	1 kV to 5 kV (50 Hz)	0.23 %	
AC Current Measure ⁴	5 kV to 28 kV (50 Hz)	2 %	Using High Voltage Divider with DMM's, Sources and HV Probe with DMM by Direct Method/Comparison Method
	28 kV to 100 kV (50 Hz)	2.3 %	
	20 μ A to 200 μ A (50 Hz to 1 kHz)	250 μ A/A + 0.02 μ A	
	200 μ A to 20 mA (50 Hz to 1 kHz)	280 μ A/A + 2 μ A	
	20 mA to 200 mA (50 Hz to 10 kHz)	250 μ A/A + 0.02 mA	
	200 mA to 2 A (50 Hz to 10 kHz)	600 μ A/A + 0.2 mA	
	2 A to 20 A (50 Hz to 2 kHz) (2 kHz to 10 kHz)	800 μ A/A + 2 mA 0.25 % + 2 mA	
1 A to 30 A (50 Hz to 5 kHz)	0.5 %	Using Shunt with DMM by V-I Method	
30 A to 1000 A (50 Hz)	0.6 %	Using Shunt with DMM by direct method Current coil & Clamp Meter	
1000 A to 3000 A (50 Hz)	2.5 %		

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION PROCEDURE AND/OR STANDARD EQUIPMENT USED
AC Power Measure ⁴ 1 Phase, 50 Hz @ UPF 120 V to 240 V, 0.01 A to 20 A	1.2 W to 4.8 kW	0.25 %	Using Digital Power Meter WT 210 by Direct Method
AC Power Measure ⁴ 1 Phase, 50 Hz, 0.8 PF Lead/Lag 120 V to 240 V, 0.1 A to 20 A	9.6 W to 3.8 kW	0.23 %	Using Digital Power Meter WT 210 by Direct Method
AC Power Measure ⁴ 1 Phase, 50 Hz, 0.5 PF Lead/Lag, 120 V to 240V, 0.1 A to 20 A	6 W to 2.4 kW	0.5 %	Using Digital Power Meter WT 210 by Direct Method
AC Power Measure ⁴ 1 Phase, 50 Hz, 0.2 PF Lead/Lag, 120 V to 240 V, 0.1 A to 20 A	2.4 W to 960 W	0.5 %	Using Digital Power Meter WT 210 by Direct Method
High Frequency Power Measure ⁴ (400 kHz)	100 mW to 400 W	4 %	Using Differential probe in combination with Oscilloscope by comparison method
DC Voltage Measure ⁴	50 µV to 0.5 mV	0.8 %	Direct / Comparison Method using 8½ DMM HP 3458A/Fluke 8508A
	0.5 mV to 200 mV	7.2 µV/V + 0.1 µV	
	200 mV to 2 V	4.6 µV/V + 0.4 µV	
	2 V to 20 V	4.6 µV/V + 4 µV	
	20 V to 200 V	6.7 µV/V + 0.04 mV	Using HV Divider with DMM by Direct Comparison Method
	200 V to 1000 V	6.8 µV/V + 0.5 mV	
	1 kV to 5 kV	0.20 %	Using Source & HV Probe with DMM by Direct Comparison Method
	5 kV to 40 kV	2 %	
	40 kV to 100 kV	1.7 %	
DC Current Measure ⁴	10 nA to 100 nA	0.05 %	Using Shunt with DMM V-I / Comparison Method
	100 nA to 10 µA	0.08 %	Using 8½ DMM HP 3458A FLUKE8508A, by Direct /Comparison Method
	10 µA to 200 µA	13 µA/A + 0.4 µA	
	200 µA to 20 mA	14 µA/A + 0.4 µA	
	20 mA to 200 mA	55 µA/A + 0.8 µA	
	200 mA to 2 A	0.02 % + 16 µA	
	2 A to 20 A	0.04 % + 0.4 mA	
	20 A to 75 A	0.08 %	
	75 A to 100 A	0.60 %	
	100 A to 1000 A	1.0 %	

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION PROCEDURE AND/OR STANDARD EQUIPMENT USED
DC Power Measure ⁴ (1 V to 600 V, 1 mA to 20 A)	1 mW to 10 W 10 W to 1 kW 1 kW to 12 kW	0.09 % 0.50 % 0.65 %	Using Digital Power Meter WT 210 by Direct Method
Pulse Energy Measure ⁴ (1 ms to 10 ms pulse length)	0.1 J to 360 J	2.8 %	Using oscilloscope DP07254, Differential Probe, DP-100 and P6015 for measurements. Defibrillator used as source
Capacitance Measure ⁴	(1 kHz) 1 pF to 1 μF 1 μF to 100 μF 100 μF to 100 mF	0.06 % 0.13 % 0.11 %	Using LCR Meter by Direct / Comparison Method
Inductance Measure ⁴	(1 kHz) 100 μH to 1 mH 1 mH to 10 H	0.09 % 0.07 %	Using LCR Meter by Direct / Comparison Method
AC Resistance Measure ⁴	(1 kHz) 1 Ω to 10 kΩ	0.06 %	Using LCR Meter
	(1 kHz to 100 kHz) 100 Ω to 1 kΩ	0.16 %	Using LCR Meter & calibrator
DC Resistance Measure ⁴	75 μΩ to 1 mΩ 1 mΩ to 10 mΩ 10 mΩ to 1 Ω 0.1 Ω to 1 Ω	0.60 % 0.60 % 0.035 % + 4 μΩ 0.035 % + 40 μΩ	Using micro-ohmmeter and Fluke 8508A DMM, standard resistors and shunts by VI method
	1 Ω to 2 Ω 2 Ω to 20 Ω 20 Ω to 200 Ω 200 Ω to 2 kΩ 2 kΩ to 20 kΩ 20 kΩ to 200 kΩ 200 kΩ to 2 MΩ 2 MΩ to 20 MΩ 20 MΩ to 200 MΩ 200 MΩ to 2 GΩ 2 GΩ to 20 GΩ	12 μΩ/Ω + 4 μΩ 7 μΩ/Ω + 30 μΩ 9 μΩ/Ω + 50 μΩ 9 μΩ/Ω + 0.5 mΩ 7.5 μΩ/Ω + 22 mΩ 9 μΩ/Ω + 100 mΩ 6.5 μΩ/Ω + 7 Ω 7 μΩ/Ω + 160 Ω 4 μΩ/Ω + 7 kΩ 30 μΩ/Ω + 0.28 MΩ 0.05 % + 20 MΩ	Using DMM 81/2 8508A by Direct Comparison method
	20 GΩ to 1 TΩ	1 %	Using Fluke Calibrator & 8½ DMM by VI Method/Comparison
Temperature Simulation - Measure ⁴	B Type: 600 °C to 1800 °C K Type: -200 °C to 1372 °C E Type: -200 °C to 1000 °C U Type: -200 °C to 400 °C L Type: -200 °C to 900 °C J Type: -200 °C to 1200 °C T Type: -200 °C to 400 °C N Type: -200 °C to 1300 °C R Type: 0 °C to 1750 °C S Type: 0 °C to 1750 °C	0.5 °C 0.05 °C 0.08 °C 0.07 °C 0.08 °C 0.06 °C 0.1 °C 0.07 °C 0.07 °C 0.07 °C	Using 8½-digit DMM 3458A/8508A DC mV measurement method

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION PROCEDURE AND/OR STANDARD EQUIPMENT USED
	RTD: -200 °C to 800 °C	0.02 °C	
Oscilloscope – DC Signal	1 mV to 130 V	0.08 %	Using Fluke Calibrator 5520A/5522A with 1.1 GHz option by Direct Method
Oscilloscope - Band Width @ 50 kHz Reference	50 kHz to 1 GHz	0.07 %	
Oscilloscope - Scope Amplitude, Square Wave Signal (10 Hz to 10 kHz)	1 mV to 55 Vp-p	0.12 %	
Oscilloscope, Time Marker	1 ns to 1000 s	0.00005 %	Using Rubidium standard, Fluke Calibrator 5520A/5522A with 1.1 GHz option by Direct Method
RF/Microwave and Electromagnetics			
3 dB Bandwidth Measure ⁴ (Filter, Power Meter, Power Sensor)	Up to 40 GHz	1 %	Using RF Reference Source - 9640ALPNX, Signal Generator and Power Meter by Direct Method
RF Power Measure ⁴ / Generate ³	(1 kHz to 18 GHz) -60 dBm to -100 dBm -60 dBm to 15 dBm	0.65 dB 0.24 dB	Using RF Reference source 9640A LPNX, Signal Generator, Attenuator, Multimeter, Power Meter, Spectrum Analyzer Method
	(18 GHz to 40 GHz) 15 dBm to -60 dBm	11 %	Using Signal Generator SMB 100A & USB Power Sensor U2054XA
	(18 GHz to 29.99 GHz) 15 dBm to -100 dBm	12 %	Using Signal Generator SMB 100A & Spectrum analyzer FSV30
RF Attenuation/Insertion Loss/Transmission Loss Measure ⁴ / Generate ³	(1 kHz) 1 dB to 60 dB	0.09 dB	Using RF Reference Source 9640A, Signal Generator-Multimeter & Power Meter Method
	(1 kHz to 18 GHz) 1 dB to 60 dB	0.14 dB	
	(18 GHz to 40 GHz) 1 dB to 60 dB	7.4 %	
	(10 MHz to 18 GHz) 60 dB to 110 dB	0.50 dB	Using Signal Generator, attenuator & Spectrum Analyzer Method
	(18 GHz to 29.99 GHz) 1 dB to 110 dB	7.4 %	

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (\pm)	CALIBRATION PROCEDURE AND/OR STANDARD EQUIPMENT USED
Spectral Purity (THD) Measure ⁴	0.17 % to 3.19 % (No. of harmonics n=2 to n=10, 10 Hz to 2.9 GHz)	0.2 %	Using Spectrum Analyzer FSV-30 Up to 30 GHz by Direct Method
Reflection Coefficient Measure ⁴ / Generate ³ - 10 MHz to 18 GHz (Maury Microwave Mismatch Test Set/Network Analyzer)	$\rho = 0.024$ to 0.33	0.032	Using Network Analyzer (R&S ZVB20) by direct method, with Cal kit Z270. Maury Microwave Mismatch Test Set used for verification of Network Analyzer performance as check/transfer standard.
Amplitude Modulation (AM) Measure ⁴ / Generate ³ CW: 100 kHz to 3.9 GHz Modulation Rate: 50 Hz to 10 kHz AM Depth	1 % to 98 %	0.25 %	Using Rhode & Schwarz Signal generator, reference source, Spectrum Analyzer FSV 30, Modulation analyzer HP 8901B as transfer by Relative Sideband Amplitude Method
Frequency Modulation (FM) Measure ⁴ / Generate ³ CW: 100 kHz to 25 GHz Modulation Rate: 50 Hz to 267 kHz FM Deviation	50 Hz to 4 MHz	0.15 %	Using Rhode & Schwarz Signal generator, reference source, Spectrum Analyzer FSV 30, by Bessel Function Method
Time/Frequency			
Frequency Generate ³	1 mHz to 40 GHz	0.2 nHz/Hz + 10 μ Hz	Using Rubidium Frequency Standard, Signal Generator, RF Reference Source by Direct Method
Frequency Measure ⁴	1 mHz to 29.999 GHz	0.2 nHz/Hz + 10 μ Hz	Using Rubidium Frequency Standard locked to other Equipment, Frequency Counter 5350B, Spectrum Analyzer -- by Direct Method
Time Interval	100 ms to 86400 s	12 μ s/s + 10 μ s	Using Timer/counter and oscilloscope by Direct method
Rotational Speed - Non-Contact	1 rpm to 100 rpm	0.58 %	Using Function generator
	100 rpm to 1000 rpm	0.23 %	Using Digital Tachometer as per SANAS TR 45
	1001 rpm to 5000 rpm	0.09 %	
	5001 rpm to 20000 rpm	0.06 %	
20001 rpm to 90000 rpm	0.06 %		

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION PROCEDURE AND/OR STANDARD EQUIPMENT USED
Rotational Speed - Contact	6 rpm to 10 rpm 10 rpm to 100 rpm 101 rpm to 4000 rpm 4001 rpm to 8000 rpm	0.4 % 0.23 % 0.09 % 0.07 %	Using Digital Tachometer as per SANAS TR 45
Rotational Speed - Centrifuge	6 rpm 7 rpm to 1000 rpm 1001 rpm to 5000 rpm 5001 rpm to 20000 rpm	0.4 % 0.18 % 0.09 % 0.07 %	Using Digital Tachometer as per SANAS TR 45
Chemical			
pH Meter	4 pH, 7 pH, 10 pH	0.08 pH	Using pH buffer solutions, By Direct Method.
Conductivity Meter	1.3 µS/cm, 5 µmS/cm, 0.015 mS/cm, 0.147 mS/cm, 1.41 mS/cm, 12.8 mS/cm, 111 mS/cm	1.4 %	Using Conductivity Solutions, By Direct Method
Optical Radiation			
Illuminance meter	1 lux to 20,000 lux	3.4 %	Using Standard Illuminance Meter, by Comparison Method
OT Lights (Light Intensity Measurement)	1 lux to 200,000 lux	3.3 %	Using Illuminance Meter by Direct Method

¹The uncertainty covered by the Calibration and Measurement Capability (CMC) is expressed as the expanded uncertainty having a coverage probability of approximately 95 %. It is the smallest measurement uncertainty that a laboratory can achieve within its scope of accreditation when performing calibrations of a best existing device. The measurement uncertainty reported on a calibration certificate may be greater than that provided in the CMC due to the behavior of the calibration item and other factors that may contribute to the uncertainty of a specific calibration.

²When uncertainty is stated in relative terms (such as percent, a multiplier expressed as a decimal fraction or in scientific notation), it is in relation to instrument reading or instrument output, as appropriate, unless otherwise indicated.

³Capability is suitable for the calibration of measuring devices in the stated ranges.

⁴Capability is suitable for the calibration of devices intended to generate the indicated quantity in the stated ranges.

SG = specific gravity
PF = power factor
THD = total harmonic distortion
CW = continuous wave